

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 05/31/2021

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:48
INTERNAL STANDARD	VERIFIED	11:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:48
BLANK TEST	0.00	11:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52

[Handwritten signature]
[Handwritten initials]

