

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 06/07/2021

OPERATOR NAME:
 PHILIP E MISSERE
 Permit #: 41131
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:12
INTERNAL STANDARD	VERIFIED	11:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:12
BLANK TEST	0.00	11:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:13
BLANK TEST	0.00	11:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:15
BLANK TEST	0.00	11:16

