

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 07/06/2021

OPERATOR NAME:
 DILLAN L CHERY
 Permit #: 42470
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 11:23 |
| INTERNAL STANDARD | VERIFIED | 11:23 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:24 |
| BLANK TEST | 0.00 | 11:25 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:25 |
| BLANK TEST | 0.00 | 11:26 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 11:26 |
| BLANK TEST | 0.00 | 11:27 |

