

SUPERVISOR TEST

-----  
 SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106  
 -----

Date: 08/16/2021

OPERATOR NAME:  
 MATTHEW J BOWE  
 Permit #: 40643  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:30
INTERNAL STANDARD	VERIFIED	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:31
BLANK TEST	0.00	09:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:34

*P.O.*  
*[Handwritten signature]*  
*[Handwritten signature]*

