

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 08/17/2021

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:51
INTERNAL STANDARD	VERIFIED	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:51
BLANK TEST	0.00	11:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:53
BLANK TEST	0.00	11:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:54
BLANK TEST	0.00	11:55

