

SUPERVISOR TEST

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 08/30/2021

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:20
INTERNAL STANDARD	VERIFIED	08:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:22
BLANK TEST	0.00	08:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:23
BLANK TEST	0.00	08:24

P.O.



