

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 09/13/2021

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 08:04 |
| INTERNAL STANDARD | VERIFIED | 08:04 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 08:04 |
| BLANK TEST | 0.00 | 08:05 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:05 |
| BLANK TEST | 0.00 | 08:06 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 08:07 |
| BLANK TEST | 0.00 | 08:07 |

P.O.
MJB
#620

