

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 09/27/2021

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:34
INTERNAL STANDARD	VERIFIED	08:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:34
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:38

Matthew J. Bowe

