

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 10/04/2021

OPERATOR NAME:
MATTHEW J BOWE
Permit #: 40643
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:47
INTERNAL STANDARD	VERIFIED	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:50
BLANK TEST	0.00	08:51

P.O. M J Bowe #620

