

SUPERVISOR TEST




 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 10/18/2021

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:27
INTERNAL STANDARD	VERIFIED	11:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:28
BLANK TEST	0.00	11:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:29
BLANK TEST	0.00	11:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:30
BLANK TEST	0.00	11:31



614