

SUPERVISOR TEST

-----  
 SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106  
 -----

Date: 11/22/2021

OPERATOR NAME:  
 ANTHONY F TEJADA  
 Permit #: 45549  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:10
INTERNAL STANDARD	VERIFIED	19:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:12
BLANK TEST	0.00	19:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:13
BLANK TEST	0.00	19:14

