

SUPERVISOR TEST

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 12/28/2021

OPERATOR NAME:  
MATTHEW J BOWE  
Permit #: 40643  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:27
INTERNAL STANDARD	VERIFIED	07:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31

Sgt. *M J Bowe*

