

SUPERVISOR TEST

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 01/10/2022

OPERATOR NAME:
 ELYOKEM G UNGAR
 Permit #: 45754
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:35
INTERNAL STANDARD	VERIFIED	23:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:35
BLANK TEST	0.00	23:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:36
BLANK TEST	0.00	23:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39

