

SOLUTION CHANGE

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 01/31/2022

OPERATOR NAME:
 ANTHONY F TEJADA
 Permit #: 45549
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:19
INTERNAL STANDARD	VERIFIED	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:22
BLANK TEST	0.00	10:23

ALCOHOL REFERENCE SOLUTION
 FOR SIMULATOR

366
 BOT. NO.

5/17/22
 EXP. DATE

21440
 LOT NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH
 A PROPERLY OPERATING BREATH SIMULATOR AND
 BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
 PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
 HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

PO 45549



