

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 02/23/2022

OPERATOR NAME:
DILLAN L CHERY
Permit #: 42470
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:12
BLANK TEST	0.00	14:13

