

SUPERVISOR TEST




SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 04/25/2022

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:23
INTERNAL STANDARD	VERIFIED	16:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:23
BLANK TEST	0.00	16:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:25
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:26
BLANK TEST	0.00	16:27



6/4