

SUPERVISOR TEST




SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 05/23/2022

OPERATOR NAME:
ANDREW S COLE-HATCHARD
Permit #: 47565
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:44
INTERNAL STANDARD	VERIFIED	17:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:44
BLANK TEST	0.00	17:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:46
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:47
BLANK TEST	0.00	17:48



② 665