

SUPERVISOR TEST




-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 07/13/2022

OPERATOR NAME:  
BRANDON KING  
Permit #: 36388  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:26
INTERNAL STANDARD	VERIFIED	08:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:27
BLANK TEST	0.00	08:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:30



*Handwritten signature/initials: BK 6/14*