

SUPERVISOR TEST




SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 07/15/2022

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:41
INTERNAL STANDARD	VERIFIED	08:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:42
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:44
BLANK TEST	0.00	08:45



Handwritten signature/initials