

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 07/23/2022

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:44
INTERNAL STANDARD	VERIFIED	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48

614

