

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106


Date: 08/01/2022

OPERATOR NAME:
 ANDREW S COLE-HATCHARD
 Permit #: 47565
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:30
INTERNAL STANDARD	VERIFIED	11:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34

PO L-Att 645



**ALCOHOL REFERENCE SOLUTION
 FOR SIMULATOR**

22190
 LOT NO.
 11/30/22
 EXP. DATE
 1182
 BOT. NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH
 A PROPERLY OPERATING BREATH SIMULATOR AND
 BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
 PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
 HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06