

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 11/03/2022

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:23
INTERNAL STANDARD	VERIFIED	18:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:23
BLANK TEST	0.00	18:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26

ALCOHOL REFERENCE SOLUTION
FOR SIMULATOR

22260
LOT NO.

1/25/23
EXP. DATE

1176
BOT. NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH
A PROPERLY OPERATING BREATH SIMULATOR AND
BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

614



