

SUPERVISOR TEST




-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 01/11/2023

OPERATOR NAME:  
BRANDON KING  
Permit #: 36388  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:22
INTERNAL STANDARD	VERIFIED	08:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:25
BLANK TEST	0.00	08:26



0614