

SUPERVISOR TEST




SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 01/16/2023

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:17
INTERNAL STANDARD	VERIFIED	16:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:18
BLANK TEST	0.00	16:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:19
BLANK TEST	0.00	16:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:20
BLANK TEST	0.00	16:21



614