

SUPERVISOR TEST




-----  
SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106  
-----

Date: 05/09/2023

OPERATOR NAME:  
BRANDON KING  
Permit #: 36388  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:14
INTERNAL STANDARD	VERIFIED	08:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18



614