

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 06/03/2023

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:29
INTERNAL STANDARD	VERIFIED	16:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:29
BLANK TEST	0.00	16:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:30
BLANK TEST	0.00	16:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:32
BLANK TEST	0.00	16:33

614

ALCOHOL REFERENCE SOLUTION
 FOR SIMULATOR

1233
 BOT. NO.

9/27/23
 EXP. DATE

23090
 LOT NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH
 A PROPERLY OPERATING BREATH SIMULATOR AND
 BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
 PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
 HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

