

SOLUTION CHANGE

 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 09/01/2023

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:17
INTERNAL STANDARD	VERIFIED	17:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:17
BLANK TEST	0.00	17:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:19
BLANK TEST	0.00	17:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:20
BLANK TEST	0.00	17:21

614

**ALCOHOL REFERENCE SOLUTION
FOR SIMULATOR**

<u>23190</u> LOT NO.	<u>11/30/23</u> EXP. DATE	<u>1044</u> BOT. NO.
-------------------------	------------------------------	-------------------------

WHEN THIS REFERENCE SOLUTION IS USED WITH
 A PROPERLY OPERATING BREATH SIMULATOR AND
 BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
 PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
 HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

