

SUPERVISOR TEST




SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 09/09/2023

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:44
INTERNAL STANDARD	VERIFIED	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:47
BLANK TEST	0.00	08:48



2614