

SUPERVISOR TEST

-----  
 SPRING VALLEY VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 127106  
 -----

Date: 09/21/2023

OPERATOR NAME:  
 ANDREW S COLE-HATCHARD  
 Permit #: 47565  
 Operator Agency/Dept: SPRING VALLEY PD  
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:28
INTERNAL STANDARD	VERIFIED	09:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:29
BLANK TEST	0.00	09:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:30
BLANK TEST	0.00	09:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:31
BLANK TEST	0.00	09:32

