

SOLUTION CHANGE

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 10/01/2023

OPERATOR NAME:
BRANDON KING
Permit #: 36388
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:32
INTERNAL STANDARD	VERIFIED	02:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:32
BLANK TEST	0.00	02:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:36

300.16
746
ELM

REFERENCE
FOR SIMULATC

1/4/24
EXP. DATE

NO.

WHEN THIS REFERENCE SOLUTION IS USED, WITH PROPERLY OPERATING I REATH SIMULATOR AND BATH ALCOHOL ANALYZER, THE SOLUTION SHOULD PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY: GUTH LABORATORIES, P.
HARRISBURG, PENNSYLVANIA 17111

Rev. 1.06

