

SUPERVISOR TEST


 SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 10/22/2023

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:36
INTERNAL STANDARD	VERIFIED	16:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:37
BLANK TEST	0.00	16:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:39
BLANK TEST	0.00	16:40



@614