

SOLUTION CHANGE

SPRING VALLEY VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 127106

Date: 12/09/2023

OPERATOR NAME:  
BRANDON KING  
Permit #: 36388  
Operator Agency/Dept: SPRING VALLEY PD  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:27
INTERNAL STANDARD	VERIFIED	18:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:27
BLANK TEST	0.00	18:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:29
BLANK TEST	0.00	18:30

ALCOHOL REFERENCE SOLUTION  
FOR SIMULATOR

601  
BOT. NO.

3/20/24  
EXP. DATE

23350  
LOT NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH  
A PROPERLY OPERATING BREATH SIMULATOR AND  
BREATH ALCOHOL ANALYZER, THE SOLUTION WILL  
PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.  
HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06

2614

