

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106


Date: 01/03/2024

OPERATOR NAME:
 MATTHEW J BOWE
 Permit #: 40643
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:39
INTERNAL STANDARD	VERIFIED	04:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:40
BLANK TEST	0.00	04:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	04:42
BLANK TEST	0.00	04:43

Sgt. M R #618



ALCOHOL REFERENCE SOLUTION
 FOR SIMULATOR

23350
 LOT NO.

3/20/24
 EXP. DATE

499
 BOT. NO.

WHEN THIS REFERENCE SOLUTION IS USED WITH
 A PROPERLY OPERATING BREATH SIMULATOR AND
 BREATH ALCOHOL ANALYZER, THE SOLUTION WILL
 PROVIDE A VALUE OF 0.10%.

PREPARED AND PACKAGED BY GUTH LABORATORIES, INC.
 HARRISBURG, PENNSYLVANIA 17111

Rev. 1/06