

SUPERVISOR TEST

SPRING VALLEY VILLAGE
POLICE DEPT
DATAMASTER dmt: 127106

Date: 02/20/2024

OPERATOR NAME:
ANTHONY F TEJADA
Permit #: 45549
Operator Agency/Dept: SPRING VALLEY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:46
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:48
BLANK TEST	0.00	22:49

