

SOLUTION CHANGE

SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 04/01/2024

OPERATOR NAME:
 DILLAN L CHERY
 Permit #: 42470
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:04
INTERNAL STANDARD	VERIFIED	23:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:04
BLANK TEST	0.00	23:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:06
BLANK TEST	0.00	23:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:07
BLANK TEST	0.00	23:08

(L)

