

SUPERVISOR TEST




SPRING VALLEY VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 127106

Date: 05/22/2024

OPERATOR NAME:
 BRANDON KING
 Permit #: 36388
 Operator Agency/Dept: SPRING VALLEY PD
 Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:16
INTERNAL STANDARD	VERIFIED	01:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20



Q614